

เลขที่ บด. 059/60

วันที่ 25 เมษายน 2560

ทนร.
เลขที่รับ..... 629
วันที่..... 27 เม.ย. 60
เวลา..... 9:44 น.

เรื่อง อนุญาตให้เปิดเผยผลการทดสอบ

เรียน ผู้จัดการส่วนบริการทดสอบและสอบเทียบ

ตามที่ บริษัท บางกอกดาต้าคอม จำกัด ได้เข้าร่วมตรวจสอบ/ทดสอบ คุณสมบัติเบื้องต้นของ เครื่องมือวัด OTDR ยี่ห้อ VeEX รุ่น FX300 กับ ส่วนบริการทดสอบและสอบเทียบ บมจ.ทีโอที ไปนั้น

ในนามของบริษัท บางกอกดาต้าคอม จำกัด ยินดีให้เปิดเผยผลการทดสอบดังกล่าว เพื่อใช้ประโยชน์ ใน กิจการของ บมจ.ทีโอที โดยไม่ขอเรียกร้องสิทธิใดๆ ทั้งสิ้นในภายหลัง

จึงเรียนมาเพื่อโปรดทราบ



ขอแสดงความนับถือ

Wara P. Pimmanongkulhan
(นายวรัทม์ ประพินมงคลการ)

กรรมการผู้จัดการ/ผู้รับมอบอำนาจ

หมายเหตุ

ทีมงานตลาด ชื่อ นางสาวสุนารี ชาวแก้ว โทร 098-258-6267, 02-645-0286-8

E-mail : sunaree_k@bdc.co.th



TEST REPORT

Requested by Bangkok Datacom Co., Ltd.

Equipment Under Test Optical Time Domain Reflectometer Model / Type FX300

Brand name/Manufacturer VeEX

Reference (Standard / Specification) No. Follow manufacturer specification

SUMMARY

Item No.	Equipment Under Test	Lot Size	Sample Size
	Optical Time Domain Reflectometer Band VeEX, Model FX300, S/N. TF3E00PD511005	1	1
	Remark 1 Can see test instruction condition from attend file - VeEX VePAL FX300		

Head of Testing Lab.
(Naritsomjarern Sumpaopol)
03 February 2017

TM
(Mr. Chairit Sridawong)
03 February 2017

Purchase Order No.	-	Contract No.	-
Document Control Code	-		



TEST REPORT

Requested by Bangkok Datacom Co., Ltd.

Equipment Under Test Optical Time Domain Reflectometer

Model / Type FX300

Brand name/Manufacturer VeEX

Reference (Standard / Specification) No.

Follow manufacturer specification

TEST DATA

Specification Descriptions	Specific	Results	Remark
1. Dynamic Range Accuracy Wavelength 1650nm (Follow "VeEX VePAL FX300" Document, Note. 2)	-	39.28 dB	
2. Event Dead Zone Accuracy Wavelength 1650nm (Follow "VeEX VePAL FX300" Document, Note. 3)	-	0.77 m	
3. Attenuation Dead Zone Accuracy Wavelength 1650nm (PW 3us, Dist. 6km and Reflection -55dB refer to VeEX spec Rev.2)	-	3.26 m	
4. Battery Life Time (Tested under standard procedure, Telcordia GR-196)	-	9.30 hr.	
5. FTTx Simulation Network Issue 1		Satisfied	
6. FTTx Simulation Network Issue 2		Satisfied	
7. FTTx Simulation Network Issue 3		Satisfied	
8. FTTx Simulation Network Issue 4		Satisfied	
(Item 5 - 8, Follow "TOT OIC Lab. Ftx Simulation Network" Document)			

Supervisor Chanathsithth T.
(Mr.Chanathsithth Traison)

Tester Knt
(Mr.Krit Srivilas)

03 February 2017

03 February 2017

Purchase Order No. _____	Contract No. _____
Document Control Code _____	